

JVC KENWOOD

JVC KENWOOD Corporation
3-12, Moriya-cho, Kanagawa-ku, Yokohama-shi,
Kanagawa, 221-0022 Japan

Declaration of Conformity

Application of council Directive---- 93/42/EEC + Amd.5 (2007/47/EC), 2011/65/EU

Standard(s) to which conformity is declared

SAFETY-----EN 60601-1: 2006 + CORR: 2010

UL Japan, Inc.

4383-326 Asama-cho, Ise-shi, Mie, 516-0021, Japan

EMC-----EN60601-1-2:2007

AUDIX Technology Corporation

EMC Department

No.53-11, Dingfu, Linkou Dist., New Taipei City 244, Taiwan

RoHS----- EN50581:2012

Others-----EN62366:2008

EN60601-1-6:2007/AC:2010

EN62304:2006

Classification-----Class I (according to Annex IX to Council Directive 93/42/EEC)

Manufacture's Name-----JVC KENWOOD Corporation.

Manufacture's Address-----3-12, Moriya-cho, Kanagawa-ku, Yokohama-shi,
Kanagawa, 221-0022 Japan

Type of Equipment-----Medical Electrical Equipment LCD Monitor

Model Number-----ML21055-#####

Brand Name-----**TOTOKU** MS55i2

Authorized Representative

Rein Medical GmbH

Schwalmstraße 301, 41238 Mönchengladbach, Germany

The model above, "ML21055-B0F0J", "ML21055-B0F0U", "ML21055-B0F0E", "ML21055-B0F0T", "ML21055-B1F0J", "ML21055-B1F0U", "ML21055-B1F0E", "ML21055-B2F0J", "ML21055-B2F0U", "ML21055-B2F0E", and "ML21055-B2F0T" are identical to the model, "ML21055-#####" listed in the test report in every aspect of attributes. The model number above has been used only for the marketing purpose.

I, the undersigned hereby declare that the equipment specified above conforms to the above Directive(s) and Standards.

Full Name Tsukasa Tashiro
Title General Manager

Signature 
Date April 7, 2014